

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 063**

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

Location(s) where activities are performed under accreditation

Head Office

Vijzelmolenlaan 7
 3447 GX
 Woerden
 The Netherlands

Location	Abbreviation/ location code
Vijzelmolenlaan 7 3447 GX Woerden The Netherlands	WO
On-site	OS

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
LF 0 0	DC/LF ELECTRICITY				
LF 1 0	DIRECT VOLTAGE				WO
	0 mV – 2 mV		$6 \cdot 10^{-6} \cdot U + 1.1 \mu\text{V}$	Generating	
	2 mV – 20 mV		$8 \cdot 10^{-6} \cdot U + 1.1 \mu\text{V}$		
	20 mV – 200 mV		$1.0 \cdot 10^{-5} \cdot U + 1.0 \mu\text{V}$		
	200 mV – 2 V		$1.1 \cdot 10^{-5} \cdot U + 1.0 \mu\text{V}$		
	2 V – 20 V		$6 \cdot 10^{-6} \cdot U + 22 \mu\text{V}$		
	20 V – 200 V		$8 \cdot 10^{-6} \cdot U + 0.2 \text{ mV}$		

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	200 V – 1000 V		$9 \cdot 10^{-6} \cdot U + 2.7 \text{ mV}$		
	0 mV – 200 mV		$3 \cdot 10^{-5} \cdot U + 0.8 \text{ } \mu\text{V}$	Measuring	
	200 mV – 2 V		$7 \cdot 10^{-6} \cdot U + 2.0 \text{ } \mu\text{V}$		
	2 V – 20 V		$7 \cdot 10^{-6} \cdot U + 20 \text{ } \mu\text{V}$		
	20 V – 200 V		$1.2 \cdot 10^{-5} \cdot U + 0.15 \text{ mV}$		
	200 V – 1000 V		$1.1 \cdot 10^{-5} \cdot U + 1.8 \text{ mV}$		
LF 2 0	DIRECT CURRENT				WO
	0 μA – 200 μA		$5 \cdot 10^{-3} \cdot I + 5 \text{ nA}$	Generating	
	200 μA – 2 mA		$5 \cdot 10^{-4} \cdot I + 12 \text{ nA}$		
	2 mA – 20 mA		$7 \cdot 10^{-5} \cdot I + 0.12 \text{ } \mu\text{A}$		
	20 mA – 200 mA		$5 \cdot 10^{-5} \cdot I + 1.8 \text{ } \mu\text{A}$		
	200 mA – 1 A		$1.2 \cdot 10^{-4} \cdot I + 30 \text{ } \mu\text{A}$		
	1 A – 2 A		$2.4 \cdot 10^{-4} \cdot I + 0,04 \text{ mA}$		
	0 μA – 200 μA		$5 \cdot 10^{-2} \cdot I + 5 \text{ nA}$	Measuring	
	200 μA – 2 mA		$5 \cdot 10^{-3} \cdot I + 0.05 \text{ } \mu\text{A}$		
	2 mA – 20 mA		$5 \cdot 10^{-4} \cdot I + 0.5 \text{ } \mu\text{A}$		
	20 mA – 200 mA		$1.3 \cdot 10^{-4} \cdot I + 5 \text{ } \mu\text{A}$		
	200 mA – 2 A		$2.4 \cdot 10^{-4} \cdot I + 0.05 \text{ mA}$		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
LF 3 0	ALTERNATING VOLTAGE				WO
	1 mV – 2 mV	30 Hz – 3.3 kHz	$1.1 \cdot 10^{-3} \cdot U + 7 \mu\text{V}$	Generating, 2-wire	
		3.3 kHz – 10 kHz	$1.8 \cdot 10^{-3} \cdot U + 7 \mu\text{V}$		
		10 kHz – 33 kHz	$4 \cdot 10^{-3} \cdot U + 7 \mu\text{V}$		
		33 kHz – 100 kHz	$1.0 \cdot 10^{-2} \cdot U + 7 \mu\text{V}$		
	2 mV – 20 mV	30 Hz – 1 kHz	$3.2 \cdot 10^{-4} \cdot U + 7 \mu\text{V}$		
		1 kHz – 3.3 kHz	$5 \cdot 10^{-4} \cdot U + 7 \mu\text{V}$		
		3.3 kHz – 10 kHz	$1.3 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$		
		10 kHz – 33 kHz	$3.5 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$		
		33 kHz – 100 kHz	$1.2 \cdot 10^{-2} \cdot U + 6 \mu\text{V}$		
	20 mV – 200 mV	30 Hz – 330 Hz	$2.3 \cdot 10^{-4} \cdot U + 12 \mu\text{V}$		
		330 Hz – 1 kHz	$3.1 \cdot 10^{-4} \cdot U + 10 \mu\text{V}$		
		1 kHz – 3.3 kHz	$6 \cdot 10^{-4} \cdot U + 6 \mu\text{V}$		
		3.3 kHz – 10 kHz	$1.3 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$		
		10 kHz – 33 kHz	$3.6 \cdot 10^{-3} \cdot U + 6 \mu\text{V}$		
		33 kHz – 100 kHz	$1.2 \cdot 10^{-2} \cdot U + 6 \mu\text{V}$		
	90 mV – 2V	10 Hz – 32 Hz	$1.3 \cdot 10^{-4} \cdot U + 50 \mu\text{V}$	Generating, 4-wire	
		32 Hz – 330 Hz	$9 \cdot 10^{-5} \cdot U + 50 \mu\text{V}$		
		330 Hz – 3.3 kHz	$6 \cdot 10^{-5} \cdot U + 30 \mu\text{V}$		
		3.3 kHz – 33 kHz	$9 \cdot 10^{-5} \cdot U + 25 \mu\text{V}$		
		33 kHz – 100 kHz	$1.0 \cdot 10^{-4} \cdot U + 0.22 \text{ mV}$		
		100 kHz – 330 kHz	$1.6 \cdot 10^{-3} \cdot U + 0.8 \text{ mV}$		
		330 kHz – 1 MHz	$1.4 \cdot 10^{-2} \cdot U + 2.0 \text{ mV}$		
	2 V – 20 V	10 Hz – 32 Hz	$1.1 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		
		32 Hz – 330 Hz	$7 \cdot 10^{-5} \cdot U + 0.5 \text{ mV}$		
		330 Hz – 33 kHz	$6 \cdot 10^{-5} \cdot U + 0.4 \text{ mV}$		
		33 kHz – 100 kHz	$2.2 \cdot 10^{-4} \cdot U + 2.0 \text{ mV}$		
	2 V – 20 V	100 kHz – 330 kHz	$1.6 \cdot 10^{-3} \cdot U + 7 \text{ mV}$		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
		330 kHz – 1 MHz	$1.1 \cdot 10^{-2} \cdot U + 12 \text{ mV}$		
	20 V – 200 V	10 Hz – 32 Hz	$1.7 \cdot 10^{-4} \cdot U + 9 \text{ mV}$		
		32 Hz – 330 Hz	$1.2 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		330 Hz – 10 kHz	$7 \cdot 10^{-5} \cdot U + 5 \text{ mV}$		
		10 kHz – 33 kHz	$8 \cdot 10^{-5} \cdot U + 6 \text{ mV}$		
		33 kHz – 100 kHz	$4 \cdot 10^{-4} \cdot U + 20 \text{ mV}$		
	200 V – 1000 V	50 Hz – 330 Hz	$9 \cdot 10^{-4} \cdot U + 50 \text{ mV}$		
		330 Hz – 10 kHz	$7 \cdot 10^{-4} \cdot U + 40 \text{ mV}$		
		10 kHz – 33 kHz	$9 \cdot 10^{-4} \cdot U + 50 \text{ mV}$		
	2 mV – 200 mV	20 Hz – 40 Hz	$3 \cdot 10^{-4} \cdot U + 15 \text{ } \mu\text{V}$	Measuring	
		40 Hz – 2 kHz	$2.8 \cdot 10^{-4} \cdot U + 15 \text{ } \mu\text{V}$		
		2 kHz – 10 kHz	$2.7 \cdot 10^{-4} \cdot U + 15 \text{ } \mu\text{V}$		
		10 kHz – 30 kHz	$5 \cdot 10^{-4} \cdot U + 20 \text{ } \mu\text{V}$		
		30 kHz – 100 kHz	$1.0 \cdot 10^{-3} \cdot U + 40 \text{ } \mu\text{V}$		
	200 mV – 2 V	20 Hz – 40 Hz	$2.1 \cdot 10^{-4} \cdot U + 55 \text{ } \mu\text{V}$		
		40 Hz – 100 Hz	$1.9 \cdot 10^{-4} \cdot U + 55 \text{ } \mu\text{V}$		
		100 Hz – 300 Hz	$1.7 \cdot 10^{-4} \cdot U + 55 \text{ } \mu\text{V}$		
		300 Hz – 1000 Hz	$1.6 \cdot 10^{-4} \cdot U + 40 \text{ } \mu\text{V}$		
		1 kHz – 3 kHz	$1.8 \cdot 10^{-4} \cdot U + 40 \text{ } \mu\text{V}$		
		3 kHz – 10 kHz	$3 \cdot 10^{-4} \cdot U + 0.05 \text{ mV}$		
		10 kHz – 60 kHz	$6 \cdot 10^{-4} \cdot U + 0.30 \text{ mV}$		
		60 kHz – 100 kHz	$6 \cdot 10^{-4} \cdot U + 0.30 \text{ mV}$		
		100 kHz – 300 kHz	$4 \cdot 10^{-3} \cdot U + 2.5 \text{ mV}$		
		300 kHz – 1 MHz	$1.2 \cdot 10^{-2} \cdot U + 24 \text{ mV}$		
	2 V – 20 V	20 Hz – 40 Hz	$2.1 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		
		40 Hz – 100 Hz	$1.9 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		
	2 V – 20 V	100 Hz – 3 kHz	$1.9 \cdot 10^{-4} \cdot U + 0.6 \text{ mV}$		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
		3 kHz – 10 kHz	$3.0 \cdot 10^{-4} \cdot U + 0.5 \text{ mV}$		
		10 kHz – 60 kHz	$6 \cdot 10^{-4} \cdot U + 3.0 \text{ mV}$		
		60 kHz – 300 kHz	$3.7 \cdot 10^{-3} \cdot U + 30 \text{ mV}$		
		300 kHz – 1 MHz	$1.2 \cdot 10^{-2} \cdot U + 0.24 \text{ V}$		
	20 V - 200 V	20 Hz – 40 Hz	$2.2 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		40 Hz – 100 Hz	$1.9 \cdot 10^{-4} \cdot U + 7 \text{ mV}$		
		100 Hz – 300 Hz	$1.7 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		300 Hz – 10 kHz	$1.8 \cdot 10^{-4} \cdot U + 5 \text{ mV}$		
		10 kHz – 30 kHz	$3.0 \cdot 10^{-4} \cdot U + 6 \text{ mV}$		
		30 kHz – 100 kHz	$7 \cdot 10^{-4} \cdot U + 30 \text{ mV}$		
	200 V - 1000 V	40 Hz – 3 kHz	$2.2 \cdot 10^{-4} \cdot U + 30 \text{ mV}$		
		3 kHz – 10 kHz	$1.9 \cdot 10^{-4} \cdot U + 0,04 \text{ V}$		
		10 kHz – 30 kHz	$4 \cdot 10^{-4} \cdot U + 0,08 \text{ V}$		
LF 4 0	ALTERNATING CURRENT				WO
	100 μ A – 200 μ A	10 Hz – 32 Hz	$3.0 \cdot 10^{-4} \cdot I + 13 \text{ nA}$	Generating	
		32 Hz – 330 Hz	$1.9 \cdot 10^{-3} \cdot I + 11 \text{ nA}$		
		330 Hz – 1000 Hz	$6 \cdot 10^{-3} \cdot I + 5 \text{ nA}$		
	200 μ A – 2 mA	10 Hz – 32 Hz	$1.6 \cdot 10^{-4} \cdot I + 0.15 \text{ } \mu\text{A}$		
		32 Hz – 330 Hz	$1.6 \cdot 10^{-4} \cdot I + 0.17 \text{ } \mu\text{A}$		
		330 Hz – 1000 Hz	$1.5 \cdot 10^{-4} \cdot I + 0.17 \text{ } \mu\text{A}$		
		1 kHz – 3.3 kHz	$2.0 \cdot 10^{-4} \cdot I + 0.20 \text{ } \mu\text{A}$		
		3.3 kHz – 5 kHz	$2.8 \cdot 10^{-4} \cdot I + 0.17 \text{ } \mu\text{A}$		
	2 mA – 20 mA	10 Hz – 32 Hz	$1.6 \cdot 10^{-4} \cdot I + 1.5 \text{ } \mu\text{A}$		
		32 Hz – 330 Hz	$1.6 \cdot 10^{-4} \cdot I + 1.7 \text{ } \mu\text{A}$		
		330 Hz – 1000 Hz	$1.4 \cdot 10^{-4} \cdot I + 1.7 \text{ } \mu\text{A}$		
	2 mA – 20 mA	1 kHz – 3.3 kHz	$2.6 \cdot 10^{-4} \cdot I + 1.7 \text{ } \mu\text{A}$		
		3.3 kHz – 5 kHz	$2.7 \cdot 10^{-4} \cdot I + 1.7 \text{ } \mu\text{A}$		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	20 mA – 200 mA	10 Hz – 32 Hz	$1.6 \cdot 10^{-4} \cdot I + 15 \mu\text{A}$		
		32 Hz – 330 Hz	$1.6 \cdot 10^{-4} \cdot I + 16 \mu\text{A}$		
		330 Hz – 1000 Hz	$1.4 \cdot 10^{-4} \cdot I + 17 \mu\text{A}$		
		1 kHz – 3.3 kHz	$2.6 \cdot 10^{-4} \cdot I + 17 \mu\text{A}$		
		3.3 kHz – 5 kHz	$2.6 \cdot 10^{-4} \cdot I + 17 \mu\text{A}$		
	200 mA – 1 A	10 Hz – 32 Hz	$4 \cdot 10^{-4} \cdot I + 0.15 \text{ mA}$		
		32 Hz – 330 Hz	$6 \cdot 10^{-4} \cdot I + 0.15 \text{ mA}$		
		330 Hz – 1000 Hz	$1.5 \cdot 10^{-3} \cdot I + 0.10 \text{ mA}$		
		1 kHz – 3.3 kHz	$5 \cdot 10^{-3} \cdot I + 0.09 \text{ mA}$		
		3.3 kHz – 5 kHz	$8 \cdot 10^{-3} \cdot I + 33 \mu\text{A}$		
	1 A - 2 A	10 Hz – 32 Hz	$8 \cdot 10^{-4} \cdot I + 0.20 \text{ mA}$		
		32 Hz – 330 Hz	$9 \cdot 10^{-4} \cdot I + 0.20 \text{ mA}$		
		330 Hz – 1000 Hz	$1.6 \cdot 10^{-3} \cdot I + 0,20 \text{ mA}$		
		1 kHz – 3.3 kHz	$5 \cdot 10^{-3} \cdot I + 0.09 \text{ mA}$		
		3.3 kHz – 5 kHz	$8 \cdot 10^{-3} \cdot I + 33 \mu\text{A}$		
	100 μA – 200 μA	50 Hz – 1000 Hz	$4 \cdot 10^{-4} \cdot I + 25 \text{ nA}$	Measuring	
		1 kHz – 5 kHz	$6 \cdot 10^{-4} \cdot I + 0,05 \mu\text{A}$		
	200 μA – 2 mA	50 Hz – 300 Hz	$4 \cdot 10^{-4} \cdot I + 0.25 \mu\text{A}$		
		300 Hz – 1000 Hz	$4 \cdot 10^{-4} \cdot I + 0.25 \mu\text{A}$		
		1 kHz – 5 kHz	$6 \cdot 10^{-4} \cdot I + 0.5 \mu\text{A}$		
	2 mA – 20 mA	50 Hz – 300 Hz	$4 \cdot 10^{-4} \cdot I + 2.5 \mu\text{A}$		
		300 Hz – 1000 Hz	$4 \cdot 10^{-4} \cdot I + 2.5 \mu\text{A}$		
		1 kHz – 5 kHz	$6 \cdot 10^{-4} \cdot I + 5 \mu\text{A}$		
	20 mA – 200 mA	50 Hz – 1000 Hz	$4 \cdot 10^{-4} \cdot I + 25 \mu\text{A}$		
	20 mA – 200 mA	1 kHz – 5 kHz	$6 \cdot 10^{-4} \cdot I + 0,05 \text{ mA}$		
	200 mA – 2 A	50 Hz – 1000 Hz	$1.2 \cdot 10^{-3} \cdot I + 0.5 \text{ mA}$		
		1 kHz – 5 kHz	$2.5 \cdot 10^{-3} \cdot I + 1.3 \text{ mA}$		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
LF 6 0	IMPEDANCE (DC/LF)				
LF 6 2	DC resistance				WO
	1 Ω		0.2 mΩ	Generating, 4-wire	
	10 Ω		0.5 mΩ		
	100 Ω		1.9 mΩ		
	1 kΩ		21 mΩ		
	10 kΩ		0.2 Ω		
	100 kΩ		2.3 Ω		
	1 MΩ		33 Ω		
	10 MΩ		0.7 kΩ		
	100 MΩ		22 kΩ		
	10 Ω		0.24 Ω	Generating, 2-wire	
	100 Ω		0.24 Ω		
	1 kΩ		0.35 Ω		
	10 kΩ		0.5 Ω		
	100 kΩ		2.6 Ω		
	1 MΩ		33 Ω		
	10 MΩ		0.7 kΩ		
	100 MΩ		22 kΩ		
	0 Ω – 20 Ω		$2.0 \cdot 10^{-5} \cdot R + 0.12 \text{ m}\Omega$	Measuring, 4-wire	
	20 Ω – 200 Ω		$1.5 \cdot 10^{-5} \cdot R + 0.5 \text{ m}\Omega$		
	200 Ω – 2000 Ω		$1.2 \cdot 10^{-5} \cdot R + 2.5 \text{ m}\Omega$		
	2 Ω – 20 kΩ		$1.2 \cdot 10^{-5} \cdot R + 25 \text{ m}\Omega$		
	20 kΩ – 200 kΩ		$2.1 \cdot 10^{-5} \cdot R + 0.4 \Omega$		
	200 kΩ – 2 MΩ		$3.1 \cdot 10^{-5} \cdot R + 10 \Omega$		
	2 MΩ – 20 MΩ		$4.0 \cdot 10^{-5} \cdot R + 0.25 \text{ k}\Omega$	Measuring, 2-wire	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	20 MΩ – 200 MΩ		$3.5 \cdot 10^{-4} \cdot R + 15 \text{ k}\Omega$		
	200 MΩ – 2 GΩ		$3.5 \cdot 10^{-3} \cdot R + 1.1 \text{ M}\Omega$		
RF 0 0	HIGH FREQUENCY QUANTITIES				
RF 2 0	Impedance				WO
	LISN Impedance	9 kHz – 30 MHz	0.3 Ω - 1.1 Ω 1.2° – 8°	50 Ω // (50 μH + 5 Ω) and 50 Ω // 50 μH	
		100 kHz – 150 MHz	0.5 Ω - 0.9 Ω 3.3° – 8°	50 Ω // (5 μH + 1 Ω) and 50 Ω // 5 μH	
	CDN Impedance	150 kHz – 300 MHz	5 Ω – 6 Ω 2.4° – 3.9°	150 Ω, 0 ° nominal	
RF 2 1	Reflection coefficient			3)	WO, OS
	Magnitude 0 – 1.0	9 kHz – 1 MHz	0.005 (0.013) + 0.007·Γ + 0.005·Γ ²	Nominal impedance 50 Ω at nominal -10 dBm RF power	
		1 MHz – 2 GHz	0.005 (0.013) + 0.003·Γ + 0.005·Γ ²		
		2 GHz – 8 GHz	0.02 (0.03) + 0.004·Γ + 0.02·Γ ²		
		8 GHz – 18 GHz	0.03 (0.04) + 0.004·Γ + 0.05·Γ ²		
	Phase -180° – +180°	9 kHz – 18 GHz	$u(\theta) = \arcsin\left(\frac{u(\Gamma)}{ \Gamma }\right)$	If the magnitude is less than its uncertainty, the phase uncertainty is ± 180°	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
RF 2 2	Transmission Coefficient			For coaxial 50 Ω devices 1, 3)	WO, OS
	Magnitude (0 – -30 dB)	9 kHz – 100 kHz	0.04 dB (0.20 dB)	at nom. -10 dBm RF power	
		100 kHz – 10 MHz	0.04 dB (0.14 dB)		
		10 MHz – 1500 MHz	0.08 dB (0.14 dB)		
		1500 MHz – 8 GHz	0.12 dB (0.18 dB)		
		8 GHz – 18 GHz	0.12 dB (0.27 dB)		
	Magnitude (-30 – -50 dB)	9 kHz – 100 kHz	0.08 dB (0.14 dB)	at nom. 0 dBm RF power	
		100 kHz – 1500 MHz	0.08 dB (0.14 dB)		
		1500 MHz – 8 GHz	0.12 dB (0.18 dB)		
		8 GHz – 18 GHz	0.15 dB (0.30 dB)		
	Magnitude (-50 – -70) dB	9 kHz – 100 kHz	0.25 dB (0.37dB)	at nom. +10 dBm * RF power (+5 dBm > 8 GHz)	
		100 kHz – 8 GHz	0.25 dB (0.31 dB)		
		8 GHz – 18 GHz	0.25 dB (0.4 dB)		
	Magnitude (-70 – -80) dB	9 kHz – 100 kHz	0.7 dB (0.9 dB)	at nom. +10 dBm * RF power (+5 dBm > 8 GHz)	
		100 kHz – 8 GHz	0.7 dB (0.8 dB)		
		8 GHz – 18 GHz	0.7 dB (0.9 dB)		
	Magnitude (-80 – -90) dB	30 kHz – 8 GHz	2 dB (2.1 dB)	at nom. +10 dBm * RF power (+5 dBm > 8 GHz)	
		8 GHz – 18 GHz	2 dB (2.2 dB)		
	Magnitude (-90 – -100) dB	30 kHz – 18 GHz	5 dB (6 dB)	at nom. +10 dBm * RF power (+5 dBm > 8GHz)	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Antenna Reflection coefficient				
	Magnitude 0 – 1.0	30 MHz – 700 MHz	$0.06 + 0.020 \cdot \Gamma + 0.008 \cdot \Gamma^2$	Nominal impedance 50 Ω at nominal -10 dBm RF power	
		700 MHz – 1500 MHz	$0.07 + 0.020 \cdot \Gamma + 0.013 \cdot \Gamma^2$		
		1500 MHz – 3000 MHz	$0.08 + 0.020 \cdot \Gamma + 0.013 \cdot \Gamma^2$		
RF 3 0	HIGH FREQUENCY POWER			1, 3)	WO
	Calibration Factors of Power Sensors	9 kHz – 10 MHz	0.05 dB	0 dBm nominal	
		10 MHz – 3 GHz	0.06 dB	0 dBm nominal	
		3 GHz – 6 GHz	0.07 dB	0 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.09 dB	0 dBm nominal	
		10 GHz – 18 GHz	0.09 dB	0 dBm nominal	
		10 MHz – 2 GHz	0.06 dB	-30 dBm nominal	
		2 GHz – 6 GHz	0.07 dB	-30 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.10 dB	-30 dBm nominal	
		10 GHz – 18 GHz	0.10 dB	-30 dBm nominal	
	Linearity of RF power				
	0 - +20 dBm	50 MHz – 500 MHz	0.05 dB		
	0 - +10 dBm	9 kHz – 10 MHz	0.13 dB		
		10 MHz – 3 GHz	0.05 dB		
		3 GHz – 6 GHz	0.08 dB		
	0 – -10 dBm	9 kHz – 10 MHz	0.13 dB		
		10 MHz – 3 GHz	0.05 dB		
		3 GHz – 6 GHz	0.05 dB		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	0 – -20 dBm	9 kHz – 10 MHz	0.26 dB		
		10 MHz – 3 GHz	0.05 dB		
		3 GHz – 6 GHz	0.08 dB		
	0 – -30 dBm	9 kHz – 10 MHz	0.26 dB		
		10 MHz- 3 GHz	0.07 dB		
		3 GHz – 6 GHz	0.08 dB		
	0 – -40 dBm	10 MHz – 6 GHz	0.09 dB		
	0 – -50 dBm	10 MHz – 6 GHz 50 MHz	0.18 dB 0,10 dB		
	0 – -60 dBm	10 MHz – 6 GHz 50 MHz	0.35 dB 0,20 dB		
	Absolute power -60 to +20 dBm			2, 3)	
		9 kHz – 10 MHz	0.06 dB	0 dBm nominal	
		10 MHz – 6 GHz	0.07 dB	0 dBm nominal	
		6 GHz – 10 GHz	0.10 dB	0 dBm nominal	
		10 GHz – 18 GHz	0.12 dB	0 dBm nominal	
	Absolute power -60 to +20 dBm			2, 3)	
		10 MHz – 6 GHz	0.07 dB	-30 dBm nominal	
		6 GHz – 10 GHz	0.10 dB	-30 dBm nominal	
		10 GHz – 18 GHz	0.14 dB	-30 dBm nominal	
	Absolute broadband power, 0 to +45 dBm				
		9 kHz – 6 GHz	0.6 dB 1.0 dB	RadiField All other DUTs	
	Frequency response of power measuring devices			2, 3)	
		9 kHz – 10 MHz	0.05 dB	0 dBm nominal	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
		10 MHz – 3 GHz	0.06 dB	0 dBm nominal	
		3 GHz – 6 GHz	0.07 dB	0 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.10 dB	0 dBm nominal	
		10 GHz – 18 GHz	0.10 dB	0 dBm nominal	
		10 MHz – 3 GHz	0.06 dB	-30 dBm nominal	
		3 GHz – 6 GHz	0.07 dB	-30 dBm nominal	
		6 GHz – 10 GHz	0.07 – 0.10 dB	-30 dBm nominal	
		10 GHz – 18 GHz	0.10 dB	-30 dBm nominal	
	EMC detectors				
	Peak, Quasi Peak, Average, RMS Sine wave accuracy	9 kHz – 0,15 MHz 0,15 MHz – 30 MHz 30 MHz – 300 MHz 300 MHz-1000 MHz 1 GHz – 18 GHz	± 0,25 dB ± 0,25 dB ± 0,4 dB ± 0,4 dB ± 0,5 – 0,8 dB	Band A, Peak, QP, AVG, RMS Band B, Peak, QP, AVG, RMS Band C, Peak, QP, AVG, RMS Band D, Peak, QP, AVG, RMS Band E, Peak, AVG, RMS	
	Peak detector, Absolute accuracy	9 kHz – 0,15 MHz 0,15 MHz – 15 MHz 15 MHz – 30 MHz 30 MHz – 300 MHz 300 MHz- 300 MHz 300MHz – 1000 MHz	± 0,5 dB ± 0,5 dB ± 0,6 dB ± 0,6 dB ± 0,4 dB ± 0,9 dB	Band A Band B Band B Band C Band D Band D	
	Peak detector, absolute calibration	1 GHz – 18 GHz	± 0,25 dB	Band E	
	Peak detector, variation with frequency	9 kHz – 1000 MHz	± 0,25 dB	Band A, B, C, D	
	Peak detector, variation with frequency	1 GHz – 18 GHz	± 0,30 dB	Band E	

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: **K 063**of **Kiwa Dare B.V.**This annex is valid from: **27-07-2022** to **01-11-2025**Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Quasi Peak detector, Response to broadband signals Absolute Calibration	9 kHz – 0,15 MHz 0,15 MHz – 30 MHz 30 MHz – 300 MHz 300 MHz 301 MHz - 650 MHz 651 MHz-1000 MHz	± 0,5 dB ± 0,5 dB ± 0,5 dB ± 0,5 dB ± 0,8 dB ± 0,9 dB	Band A Band B Band C Band D Band D Band D	
	Quasi Peak detector, variation to pulse freq Band A Band B Band C Band D	Single to 100 Hz Single to 1000 Hz Single – 1 Hz 2 Hz 10Hz – 1000 Hz Single – 1 Hz 2 Hz 10Hz – 1000 Hz	±0,5 dB ±0,5 dB ±(0,8 – 1,3) dB ±(0,6 - 1,0) dB ±(0,5 – 1,0) dB ±(1,3 to 1,8) dB ±(1,0 to 1,7) dB ±(0,5 to 0,9) dB	Band A Band B Band C Band C Band C Band D Band D Band D	
	AVG detector, Absolute accuracy	9 kHz – 0,15 MHz 0,15 MHz – 30 MHz 30 MHz – 300 MHz 300 MHz- 1000 MHz 1 GHz – 18 GHz	± 0,5 dB ± 0,5 dB ± 0,5 dB ± 0,5 dB ± 0,5 – 1,0 dB	Band A Band B Band C Band D Band E	
	AVG detector, variation with frequency	9 kHz – 0,15 MHz 0,15 MHz – 30 MHz 30 MHz – 300 MHz 300 MHz- 1000 MHz 1 GHz – 18 GHz	± 0,25 dB ± 0,25 dB ± 0,25 dB ± 0,25 dB ± 0,25 – 0,4 dB	Band A Band B Band C Band D Band E	
	AVG detector, Intermittent, unsteady, drifting narrow band	9 kHz – 18 GHz	± 0,25 dB	Band A, B, C, D, E	
	RMS detector, Absolute accuracy	9 kHz – 0,15 MHz 0,15 MHz – 30 MHz 30 MHz – 300 MHz 300 MHz- 1000 MHz 1 GHz – 18 GHz	± 0,5 dB ± 0,5 dB ± 0,5 dB ± 0,5 dB ± 0,6 – 1,2 dB	Band A Band B Band C Band D Band E	
	RMS detector, variation with frequency	9 kHz – 18 GHz	± 0,25 dB	Band A, B, C, D, E	
	RMS detector, Intermittent, unsteady, drifting narrow band	9 kHz – 18 GHz	± 0,25 dB	Band A, B, C, D, E	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Bandwidth of RBW filters				
	1 Hz – 10 MHz	9 kHz – 2.4 GHz	0.8 + 0.02·BW		
	1 Hz – 10 MHz	10 MHz – 18 GHz	0.08 + 0.02·BW		
RF 5 0	ELECTRICAL / MAGNETIC FIELD QUANTITIES /EMC				WO
	Electrical Field Strength (1 – 200) V/m	9 kHz – 30 MHz	(0.5 – 0.6) dB	Temcell 4)	
		30 MHz – 75 MHz	(0.6 – 1.3) dB		
		75 MHz – 200 MHz	1.3 dB		
	Electrical Field Strength (1 – 100) V/m	200 MHz – 1 GHz	1.2 dB	Anechoic Chamber	
		1 GHz – 8 GHz	1.1 dB		
		8 GHz – 12 GHz	1.2 dB		
		12 GHz –15 GHz	(1.2 - 1.5) dB		
		15 GHz – 18 GHz	1.5 dB		
		18 GHz – 40 GHz	2.4 dB		
	Antenna factor				
	ANSI C63.5 CISPR 16-1-6	30 MHz – 5 GHz	0.8 dB	OATS, Standard Site method 5)	
	ANSI C63.5 CISPR 16-1-6	30 MHz – 1 GHz	0.8 dB	OATS, Reference antenna method 5)	
	Antenna symmetry - Dipole - Biconical - Hybrid	30 MHz – 1 GHz	0.25 dB	OATS ANSI C63.5 CISPR 16-1-4 5)	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Quasi Free Space	20 MHz – 100 MHz	0.9 dB	Free Space Environment, Three Antenna Method 5)	
		100 MHz – 200 MHz	0.8 dB		
		200 MHz – 5 GHz	0.7 dB		
		1 GHz – 10 GHz	1.4 dB	Full Anechoic Room, Three Antenna Method 4)	
		10 GHz – 12 GHz	(1.4 – 2.1) dB		
		12 GHz – 18 GHz	2.1 dB		
	SAE ARP 958	20 MHz – 100 MHz	0.9 dB	For military or automotive use 5)	
		100 MHz – 200 MHz	0.8 dB		
		200 MHz – 5 GHz	0.7 dB		
	Shielding Effectiveness			According to EN50147 and Mil Std 285 5)	OS
	Magnetic Field 100 dB	10 kHz – 30 MHz	5 dB		
	Electric Field 120 – 150 dB	10 MHz – 300 MHz	5 dB		
	Plane wave 110 – 140 dB	30 MHz – 1 GHz	5 dB		
	Plane wave 110 – 140 dB	1 GHz – 18 GHz	6 dB		
	Normalized Site Attenuation			According to CISPR 16-1-4 using broadband antennae Horizontal and vertical polarization, distance between 3 m and 30 m 5)	OS
	NSA	30 MHz – 1000 MHz	1.6 dB		
	Site Voltage Standing Wave Ratio			According to CISPR 16-1-4 using reciprocal method 4)	OS
	S _{VSWR}	1 GHz – 18 GHz	2.0 dB		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Field Uniformity			According to IEC 61000-4-3 4)	OS
	Forward Power	80 MHz – 18 GHz	1.3 dB		
	Field Uniformity	80 MHz – 18 GHz	1.7 dB		
	Surge generators and coupling/decoupling networks waveform surge voltage			According to EN 61000-4-5 1.2/50 µs pulse 10/700 µs pulse	WO, OS
	0 V – 550 V		$6.7 V + 0.022 \cdot U$ $6.7 V + 0.025 \cdot U$	Coupling/decoupling networks for AC/DC power supply circuits only in combination with appropriate surge generator	
	0 V – 1.1 kV		$13.4 V + 0.022 \cdot U$ $13.4 V + 0.025 \cdot U$		
	0 V – 2.8 kV		$33.5 V + 0.022 \cdot U$ $33.5 V + 0.025 \cdot U$		
	0 V – 5.5 kV		$67 V + 0.022 \cdot U$ $67 V + 0.025 \cdot U$		
	Waveform surge current				WO, OS
	Current amplitude			1.2/50 µs pulse 10/700 µs pulse	
	0 – 15 A		$0.18 A + 0.022 \cdot I$ $0.18 A + 0.029 \cdot I$	Measurements at coupling/decoupling network input, output; coupling modes line to neutral, line to earth and neutral to earth	
	0 – 30 A		$0.36 A + 0.022 \cdot I$ $0.36 A + 0.029 \cdot I$		
	0 – 60 A		$0.72 A + 0.022 \cdot I$ $0.72 A + 0.029 \cdot I$		
	0 – 150 A		$1.8 A + 0.022 \cdot I$ $1.8 A + 0.029 \cdot I$		
	0 – 300 A		$3.6 A + 0.022 \cdot I$ $3.6 A + 0.029 \cdot I$		
	0 – 600 A		$7.2 A + 0.022 \cdot I$ $7.2 A + 0.029 \cdot I$		
	0 – 1500 A		$18 A + 0.022 \cdot I$ $18 A + 0.029 \cdot I$		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	0 – 3000 A		36 A + 0.022·I 36 A + 0.029·I		
	Front time Voltage mode				WO, OS
	Pulse 1.2/50 µs	0.65 µs – 1.75 µs	0.08 µs	4)	
	Pulse 10/700 µs	5.5 µs – 15.5 µs	0.5 µs		
	Front time Current mode				WO, OS
	Pulse 1.2/50 µs	6.4 µs – 9.6 µs 1.4 µs – 3.6 µs	0.18 µs 0.07 µs	line – line line – PE	
	Pulse 10/700 µs	3.5 µs – 6.5 µs	0.27 µs		
	Duration Voltage Mode			4)	WO, OS
	Pulse 1.2/50 µs	35 µs – 65 µs	1.0 µs		
	Pulse 10/700 µs	490 µs – 910 µs	14 µs		
	Duration time Current Mode			4)	WO, OS
	Pulse 1.2/50 µs	11 µs – 21 µs	0.23 µs	Current, line to line	
		14 µs – 36 µs	0.23 µs	Current, line to earth	
	Pulse 10/700 µs	210 µs – 390 µs	6 µs		
	EFT/burst generators waveform (im)pulse. voltage into 50 Ω			According to EN 61000-4-4 (July 2004)	WO, OS
	0 V – 150 V		2.0 V + 0.022·U 2.0 V + 0.025·U	4)	
	0 – 300 V		4.0 V + 0.022·U 4.0 V + 0.025·U	4)	
	0 – 600 V		8.0 V + 0.022·U 8.0 V + 0.025·U	4)	

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	0 – 1.5 kV		20 V + 0.022 · U 20 V + 0.025 · U	4)	
	0 – 3 kV		37 V + 0.022 · U 37 V + 0.025 · U	4)	
	EFT/burst generators Waveform pulse voltage into 1 kΩ				WO, OS
	0 V – 500 V		6 V + 0.045 · U 6 V + 0.048 · U	4)	
	0 V – 1 kV		12 V + 0.045 · U 12 V + 0.048 · U	4)	
	0 V – 2 kV		24 V + 0.045 · U 24 V + 0.048 · U	4)	
	0 V – 5 kV		60 V + 0.045 · U 60 V + 0.048 · U	4)	
	Rise time (10 % – 90 %) 3 ns – 7 ns		0.3 ns	4)	WO, OS
	Pulse duration time (50 % – 50 %) 30 ns – 70 ns		2.0 ns	4)	WO, OS
	Repetition time			4)	
	5 μs – 15 μs		0.15 μs		
	150 μs – 600 μs		2.5 μs		
	Burst duration			4)	WO, OS
	10 ms – 20 ms		0.5 μs + 0.0005 · t		
	Burst period			4)	WO, OS
	200 ms – 400 ms		500 μs		

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022** to **01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	ESD Simulators Waveform discharge current				
	First peak current 3 – 10 A 6 – 20 A 15 – 50 A		7 % 7 % 7 %	Standard and Networks: 150 pF / 330 Ω IEC 61000-4-2 (2008-12)	
	Rise time 0.5 – 1.2 ns		15 %		
	Current at t1 and t2 1 – 10 A 2 – 20 A 5 – 50 A		7 % 7 % 7 %		
	ESD Simulators Waveform discharge current				
	First peak current 3 – 10 A 6 – 20 A 15 – 50 A		0.08 + 0.032·I 0.16 + 0.032·I 0.4 + 0.032·I	Standard and networks: ISO 10605 (2008-07) All networks	
	Rise time 0.5 – 1.2 ns		0.08 ns		
	Current at t1 and t2 1 – 10 A 2 – 20 A 5 – 50 A		0.08 + 0.032·I 0.16 + 0.032·I 0.4 + 0.032·I	Standard and networks: ISO 10605 (2008-07) 150 pF / 330 Ω 330 pF / 330 Ω	
	Current at t1 and t2 0.1 – 1.0 A 0.2 – 2.0 A 0.5 – 5.0 A		0.008 + 0.023·I 0.02 + 0.023·I 0.04 + 0.023·I	Standard and networks: ISO 10605 (2008-07) 150 pF / 2000 Ω 330 pF / 2000 Ω	
TF 0 0	TIME AND FREQUENCY				
TF 2 1	Frequency				WO
	10 mHz – 2.7 GHz		$(8 \cdot 10^{-10} / \tau + 1.8 \cdot 10^{-10}) \cdot f$	Measuring, 10 ms ≤ τ ≤ 400 s	

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 063**

of **Kiwa Dare B.V.**

This annex is valid from: **27-07-2022 to 01-11-2025**

Replaces annex dated: **06-07-2022**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	10 mHz – 10 Hz		$3 \cdot 10^{-5} \cdot f$	Generating	
	10 Hz – 100 kHz		$(1 \cdot 10^{-5} / \tau) \cdot f$	Generating, 10 ms ≤ τ ≤ 400 s	
	100 kHz – 2.16 GHz		$(1 \cdot 10^{-9} / \tau + 1.8 \cdot 10^{-10}) \cdot f$	Generating, 10 ms ≤ τ ≤ 400 s	
TF 2 2	Time interval				WO
	0.5 ns – 10 μs		$(1.5 \cdot 10^{-4}) \cdot t + 15 \text{ ps}$	Generating	
	10 μs – 1000 s		$(3 \cdot 10^{-5}) \cdot t$		

The calibrations are carried out at an ambient temperature of (23 ± 2) °C and a relative humidity of (50 ± 10) %, with an exception for calibrations marked 4 or 5.

1. $|\Gamma_{dut}| < 0.02$
2. $|\Gamma_{dut}| < 0.2$
3. All calibrations are based on equipment using N-type connectors.
4. The calibrations are carried out at ambient conditions within (23 ± 7) °C and (50 ± 20) %.
5. The calibrations are carried out at ambient conditions within (20 ± 15) °C and (50 ± 40) %.